NOTES:

① JUNCTION BOXES: J2--EMERGENCY POWER. BOTH E4 BOXES ARE TO BE J2 WITH KNOCK OUTS ON END PLATES. WIRES TO BE CAPPED ON BOTH ENDS FOR FIELD CONNECTIONS AT EITHER END USING HEALTH CARE FACILITIES AC OR MC (HCFAC/HCFMC) CABLE.

② GAS PIPING: HOSE CONNECTED GAS OUTLETS.

③ PANEL LAMINATE COLOR: TO BE DETERMINED.

④ MEDICAL GAS OUTLETS AND ELECTRICAL DEVICES TO BE MSC STANDARD BRAND.

⑤ ALL EXPOSED METAL SURFACES ARE TO HAVE A POWDER COATED FINISH THAT MATCHES FORMICA #923 "SURF", EXCEPT FOR THE HORIZONTAL CHANNELS WHICH ARE ANODIZED ALUMINUM.

LEGEND

REQ	SYM	DESCRIPTION
2	O	OXYGEN OUTLET
1	A	AIR OUTLET
2	E4	DUPLEX RECEPTACLE, 8300 RED, 20A, 125V, H.G., EMER. POWER
2	HC	HORIZONTAL ACCESSORY CHANNEL

QTY	ACCESSORIES NOT SHOWN (PER UNIT)

MODULAR SERVICES COMPANY
500 EAST BRITTON ROAD
OKLAHOMA CITY, OK 73114

PROJECT

REVISED

PROVIDER PLUS®
SINGLE TIER
MODEL 6540-AC

SCALE: =1'

DATE	DRAWN BY	DRAWING #
		PB-223

__ REQ'D: __ AS SHOWN
__ OPPOSITE

SERVICES TO BE REVERSED
ON OPPOSITE HAND UNITS